

TRANSLATION

PATENT COOPERATION TREATY

PCT

INTERNATIONAL PRELIMINARY REPORT ON PATENTABILITY

(Chapter II of the Patent Cooperation Treaty)

(PCT Article 36 and Rule 70)

Applicant's or agent's file reference FPCT11004	FOR FURTHER ACTION	See Form PCT/IPEA/416
International application No. PCT/JP2004/016747	International filing date (<i>day/month/year</i>) 11.11.2004	Priority date (<i>day/month/year</i>) 21.11.2003
International Patent Classification (IPC) or national classification and IPC A61B8/08		
Applicant Hitachi Medical Corporation		

1. This report is the international preliminary examination report, established by this International Preliminary Examining Authority under Article 35 and transmitted to the applicant according to Article 36.
2. This REPORT consists of a total of <u>6</u> sheets, including this cover sheet.
3. This report is also accompanied by ANNEXES, comprising: a. <input type="checkbox"/> (<i>sent to the applicant and to the International Bureau</i>) a total of _____ sheets, as follows: <input type="checkbox"/> sheets of the description, claims and/or drawings which have been amended and are the basis for this report and/or sheets containing rectifications authorized by this Authority (see Rule 70.16 and Section 607 of the Administrative Instructions). <input type="checkbox"/> sheets which supersede earlier sheets, but which this Authority considers contain an amendment that goes beyond the disclosure in the international application as filed, as indicated in item 4 of Box No. I and the Supplemental Box. b. <input type="checkbox"/> (<i>sent to the International Bureau only</i>) a total of (indicate type and number of electronic carrier(s)) _____, containing a sequence listing and/or tables related thereto, in computer readable form only, as indicated in the Supplemental Box Relating to Sequence Listing (see Section 802 of the Administrative Instructions).
4. This report contains indications relating to the following items: <input checked="" type="checkbox"/> Box No. I Basis of the report <input type="checkbox"/> Box No. II Priority <input type="checkbox"/> Box No. III Non-establishment of opinion with regard to novelty, inventive step and industrial applicability <input checked="" type="checkbox"/> Box No. IV Lack of unity of invention <input checked="" type="checkbox"/> Box No. V Reasoned statement under Article 35(2) with regard to novelty, inventive step or industrial applicability; citations and explanations supporting such statement <input type="checkbox"/> Box No. VI Certain documents cited <input type="checkbox"/> Box No. VII Certain defects in the international application <input type="checkbox"/> Box No. VIII Certain observations on the international application

Date of submission of the demand	Date of completion of this report
Name and mailing address of the IPEA/JP	Authorized officer
Facsimile No.	Telephone No.

INTERNATIONAL PRELIMINARY REPORT ON PATENTABILITY

International application No.

PCT/JP2004/016747

Box No. I Basis of the report

1. With regard to the language, this report is based on the international application in the language in which it was filed, unless otherwise indicated under this item.
- ☐ This report is based on translations from the original language into the following language _____, which is the language of a translation furnished for the purposes of:
- ☐ international search (Rule 12.3 and 23.1(b))
- ☐ publication of the international application (Rule 12.4)
- ☐ international preliminary examination (Rule 55.2 and/or 55.3)
2. With regard to the elements of the international application, this report is based on *(replacement sheets which have been furnished to the receiving Office in response to an invitation under Article 14 are referred to in this report as "originally filed" and are not annexed to this report)*:
- ☒ the international application as originally filed/furnished
- ☐ the description:
- pages _____ as originally filed/furnished
- pages* _____ received by this Authority on _____
- pages* _____ received by this Authority on _____
- ☐ the claims:
- nos. _____ as originally filed/furnished
- nos.* _____ as amended (together with any statement) under Article 19
- nos.* _____ received by this Authority on _____
- nos.* _____ received by this Authority on _____
- ☐ the drawings:
- sheets _____ as originally filed/furnished
- sheets* _____ received by this Authority on _____
- sheets* _____ received by this Authority on _____
- ☐ a sequence listing and/or any related table(s) – see Supplemental Box Relating to Sequence Listing.
3. ☐ The amendments have resulted in the cancellation of:
- ☐ the description, pages _____
- ☐ the claims, nos. _____
- ☐ the drawings, sheets/figs _____
- ☐ the sequence listing (*specify*): _____
- ☐ any table(s) related to sequence listing (*specify*): _____
4. ☐ This report has been established as if (some of) the amendments annexed to this report and listed below had not been made, since they have been considered to go beyond the disclosure as filed, as indicated in the Supplemental Box (Rule 70.2(c)).
- ☐ the description, pages _____
- ☐ the claims, nos. _____
- ☐ the drawings, sheets/figs _____
- ☐ the sequence listing (*specify*): _____
- ☐ any table(s) related to sequence listing (*specify*): _____

* If item 4 applies, some or all of those sheets may be marked "superseded."

INTERNATIONAL PRELIMINARY REPORT ON PATENTABILITY

International application No.

PCT/JP2004/016747

Box No. IV

Lack of unity of invention

1. ☒ In response to the invitation to restrict or pay additional fees the applicant has:
- ☐ restricted the claims.
 - ☐ paid additional fees.
 - ☐ paid additional fees under protest.
 - ☒ neither restricted the claims nor paid additional fees.
2. ☐ This Authority found that the requirement of unity of invention is not complied with and chose, according to Rule 68.1, not to invite the applicant to restrict or pay additional fees.
3. This Authority considers that the requirement of unity of invention in accordance with Rules 13.1, 13.2 and 13.3 is:
- ☐ complied with.
 - ☒ not complied with for the following reasons:

The feature that is common to the inventions set forth in claims 1 to 12 is the configuration of the ultrasonic diagnostic device equipped with an ultrasonic probe, a tomogram composing means, an elastic image composing means and a display means. However, as a result of the search it became apparent that the abovementioned configuration is not novel in as much as said configuration is disclosed in the document Naotaka NITTA, "Sanjigen Hen'I Vector Keisoku ni Motozuku Soshiki Danseiritsu Saikosei no Jikkenteki Kento," Journal of Medical Ultrasonics, 15 April 2000, Vol. 27, No. 4, page 738.

[Refer to the Supplemental Box]

4. Consequently, this report has been established in respect of the following parts of the international application:
- ☐ all parts.
 - ☒ the parts relating to claims Nos. 1-5, 9-12

INTERNATIONAL PRELIMINARY REPORT ON PATENTABILITY

International application No.

PCT/JP2004/016747

Box No. V	Reasoned statement under Article 35(2) with regard to novelty, inventive step or industrial applicability; citations and explanations supporting such statement		
1.	Statement		
	Novelty (N)	Claims <u>3-5</u>	YES
		Claims <u>1, 2, 9-12</u>	NO
	Inventive step (IS)	Claims _____	YES
		Claims <u>1-5, 9-12</u>	NO
	Industrial applicability (IA)	Claims <u>1-5, 9-12</u>	YES
		Claims _____	NO
2.	Citations and explanations (Rule 70.7)		
	<p>Document 1: Naotaka NITTA, "Sanjigen Hen'i Vector Keisoku ni Motozuku Soshiki Danseiritsu Saikosei no Jikkenteki Kento," Journal of Medical Ultrasonics, 15 April 2000, Vol. 27, No. 4, page 738</p> <p>Document 2: JP 9-182751 A (Hewlett Packard Co.), 15 July 1997, entire text, all drawings & US 5622173 A</p> <p>Claims 1, 2 and 9 to 12</p> <p>The inventions set forth in claims 1, 2 and 9 to 12 lack novelty and do not involve an inventive step in the light of document 1 cited in the international search report. Document 1 presents an ultrasonic diagnostic device for generating elastic images, wherein the physical values (i.e. the deformations) within an elastic image are displayed by means of an adjacent color bar, which has been configured so that the scale of said color bar changes along the deformation direction.</p> <p>Claims 3 to 5</p> <p>The invention set forth in claims 3 to 5 lacks novelty and does not involve an inventive step in the</p>		

INTERNATIONAL PRELIMINARY REPORT ON PATENTABILITY

International application No.

PCT/JP2004/016747

Box No. V

Reasoned statement under Article 35(2) with regard to novelty, inventive step or industrial applicability;
citations and explanations supporting such statement

light of documents 1 and 2 cited in the international search report. Document 2 discloses an ultrasonic diagnostic device equipped with a base line knob for changing the position of the base line that divides the color bar. Such being the case, it would have been obvious to a person skilled in the art to configure so that the technical features associated with the base line from the invention disclosed in document 2 are applied in the invention that is disclosed in document 1.

Supplemental Box

In case the space in any of the preceding boxes is not sufficient.

Continuation of:

Box IV.3

Such being the case, the abovementioned configuration cannot be considered to be a special technical feature in the meaning of the second sentence of PCT Rule 13.2

Consequently, the inventions set forth in claims 1 to 12 clearly do not conform to the requirement of unity of invention.